



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
9200 Corporate Blvd.
Rockville MD 20850

Mr. Kwang C. Jung
President
UMI International
P.O. Box 170
NORWOOD NJ 07648

DEC 28 2006

Re: K062788
Trade/Device Name: Collimator (manual drive) – a beam limiting device
Regulation Number: 21 CFR 892.1610
Regulation Name: Diagnostic x-ray beam-limiting device
Regulatory Class: II
Product Code: KPW
Dated: December 4, 2006
Received: December 6, 2006

Dear Mr. Jung:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.



Protecting and Promoting Public Health

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.


This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of this letter:

21 CFR 876.xxx	(Gastroenterology/Renal/Urology)	240-276-0115
21 CFR 884.xxx	(Obstetrics/Gynecology)	240-276-0115
21 CFR 894.xxx	(Radiology)	240-276-0120
Other		240-276-0100

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure



P. O. Box 170
Norwood, NJ 07648, U.S.A.
Tel (201)784-2950 Fax (201)784-1190

Statement of Indications for Use

510(k) Number(if known) K062788
Device Name : Collimator(manual drive) – a beam limiting device
Indications for Use :

A diagnostic x-ray beam-limiting device intended to restrict the dimension of a diagnostic x-ray field by limiting the size of the primary x-ray beam coming from x-ray tube of the x-ray system.

Prescription use ✓
(Part 21.CFR 801 Subpart D)

and/or

Over the-Counter Use _____
(21 CFR 801 Subpart C)

(Please don not write below this line – Continue on another page if needed)

Concurrence of CDRH, Office of Device Evaluation(ODE)

(Posted Nov. 13, 2003)

Nancy C Brody
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K062788